



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/550,734

09/26/2005

Brian Nielsen

P70816US0

8354

136 7590 06/25/2008

JACOBSON HOLMAN PLLC
400 SEVENTH STREET N.W.
SUITE 600
WASHINGTON, DC 20004

EXAMINER

JACKSON, BRANDON LEE

ART UNIT

PAPER NUMBER

3772

MAIL DATE

DELIVERY MODE

06/25/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/550,734	Applicant(s) NIELSEN ET AL.	
	Examiner BRANDON JACKSON	Art Unit 3772	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 09 April 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 11-24 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 11-24 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

This action is in response to amendments/arguments filed 4/9/2008. Currently, claims 11-24 and pending in the instant application.

Response to Arguments

Applicant argues the Bray device could not have a reinforcing layer density less than 50 g/m². However, Bray states the preferable range is between 50 and 350 g/m², but not that it has to be in that specific density range in order to function, since it is merely preferably. Moreover, Applicant's specification states the density range of the reinforcing layer can be from 5 to 200 g/m²; therefore, the Bray device would function the same as Applicant's claimed invention and the optimum density range can be determined via testing of the device. It would be obvious to one of ordinary skill in the art at the time of the invention to have the fabric layer density to be 5 – 60 g/m² and the reinforcing layer to have a density range of 20-30 g/m²; since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, 105 USPQ 233. Obviously, combined densities in these ranges could add up to 50 g/m².

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the

Art Unit: 3772

invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148

USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Claims 11-14, 16-18, and 20 are rejected under 35 U.S.C. 103(a) as being unpatentable over Bray et al. (UK Patent Application Publication 2,377,177). Bray discloses a wound dressing (pg. 1, lines 1-2) comprising a web of gel forming fibers (pg. 1, lines 4-5) attached to a reinforcing layer (pg. 1, lines 5-6), wherein the density of the web is in a range of 25-200 grams per square meter, which has a portion that falls within Applicant's range of 5-60 grams per square meter. The gel-forming fibers comprise alginate (pg. 1, lines 6-7). The reinforced layer and the web are attached by needling (pg. 3, lines 6-7) or thermal bonding (pg. 3, lines 9-10). The reinforced layer is woven of nonwoven fabric (pg. 1, lines 17-19). The dressing comprises an active agent that is an anti-bacterial agent (pg. 1, lines 23-24). The dressing comprises silver calcium alginate or silver calcium alginate (pg. 1, lines 12-14). Bray fails to explicitly state the reinforcing layer density less than 50 g/m², preferably between 20 and 30 g/m². However, Bray states the preferable range is between 50 and 350 g/m², but not that it has to be in that specific density range in order to function, since it is merely preferably. Moreover, Applicant's specification states the density range of the reinforcing layer can be from 5

to 200 g/m²; therefore, the Bray device would function the same as Applicant's claimed invention and the optimum density range can be determined via testing of the device. It would be obvious to one of ordinary skill in the art at the time of the invention to have the fabric layer density to be 5 – 60 g/m² and the reinforcing layer to have a density range of 20-30 g/m²; since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, 105 USPQ 233. Obviously, combined densities in these ranges could add up to 50 g/m².

Claim 15 is rejected under 35 U.S.C. 103(a) as being unpatentable over Bray et al. (UK Patent Application Publication 2,377,177) in view of Sessions et al. (US Patent 6,346,653). Bray substantially discloses the claimed invention; see rejection to claim 11 above. Bray fails to disclose an adhesive means for attaching the web to the reinforcing layer. However, Sessions discloses a wound dressing (10) comprising an adhesive means (28) for attaching a first layer (30) to a second layer (32). Therefore, it would be obvious to one of ordinary skill in the art at the time of the invention to substitute Bray bonding means for the adhesive means, as taught by Sessions, because the adhesive is a well known means in the art for securing layers of a wound dressing.

Claim 19 is rejected under 35 U.S.C. 103(a) as being unpatentable over Bray et al. (UK Patent Application Publication 2,377,177) in view of Nielsen (US Patent 6,998,509). Bray substantially discloses the claimed invention, specifically including an

active ingredient within the dressing; see rejections to claims 11 and 17 above. Bray fails to disclose a pain-relieving agent incorporated in the dressing. However, Nielsen teaches a wound care device comprising a pain-relieving agent (col. 9, lines 27-31). Therefore, it would be obvious to one of ordinary skill in the art at the time of the invention to modify the active ingredient in the dressing of Bray to substitute a pain relieving agent for the active ingredient, as taught by Nielsen, in order to provide the user with the comfort of not feeling the pain from the wound under the dressing. Such a modification would have been obvious on one of ordinary skill in the art at the time of the invention was made to substitute the pain relieving agent for the antibacterial agent, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. *In re Leshin*, 125 USPQ 416.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the

shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to BRANDON JACKSON whose telephone number is (571)272-3414. The examiner can normally be reached on Monday - Friday 8-5:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Patricia Bianco can be reached on (571)272-4940. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Brandon Jackson/
Examiner, Art Unit 3772

BLJ

Application/Control Number: 10/550,734
Art Unit: 3772

Page 7

/Patricia Bianco/

Supervisory Patent Examiner, Art Unit 3772